



EU Declaration of Conformity

5500 Vibration Meter

| | |
|--|---|
| Manufacturer's Name | Metrix Instrument Co., L.P. |
| Manufacturer's Address | 8824 Fallbrook Drive, Houston, TX 77064 USA 281.940.1802 www.metrixvibration.com info@metrixvibration.com |
| Declares that the product | Model 5500 Vibration Meter (This declaration of conformity is issued under the sole responsibility of Metrix Instrument) |
| Directive 2014/34/EU ATEX | Provision of the Directive fulfilled by the Equipment <ul style="list-style-type: none">•  II 1 G, Ex ia IIC T3 Ga (Ta -10°C to +60°C) |
| Conforms to the following standards | ATEX <ul style="list-style-type: none">• EN 60079-0: 2009• EN 60079-11: 2007 |
| EU-Type Examination | Certificate <ul style="list-style-type: none">• Baseefa03ATEX0556• SGS Baseefa Limited Rockhead Business Park, Staden Lane Buxton, Derbyshire SK17 9RZ UNITED KINGDOM Notified Body Number: 1180 |
| Notified Body for production | SGS Fimko Oy (Notified Body Number: 0598) Helsinki, Finland |
| Directive 2014/30/EU | EMC <ul style="list-style-type: none">• EN 61326-1: 2013 (self-certified)• EN 61326-1: 2006• Professional Testing (EMI), Inc. 1601 North A.W. Grimes Blvd., Suite B Round Rock, Texas 78665 NVLAP: Lab Code 200062-0 |
| Technical File Location | The Technical File required by this Directive is maintained by the manufacturer listed above and the address listed above. |
| Approved By |  Lance Truong Quality Manager, Metrix Instrument Co. Houston, Texas 77064 October 23, 2020 |